

MiniCollect® Capillary Blood Collection System

Item No.	Description	Cap Colour	Fill Vol.	Area of Application
MiniCollect® Tubes				
450412	Z No Additive	white	---	For example: CSF determinations
450447	9NC Coagulation 3.2% (liquid dose)	light blue	0.5ml	Coagulation tests (only suitable for venous blood)
450413	9NC Coagulation 3.2% (liquid dose)	light blue	1ml	Coagulation tests (only suitable for venous blood)
450409	9NC Coagulation 3.8% (liquid dose)	light blue	1ml	Coagulation tests (only suitable for venous blood)
450471	Z Serum	red	0.5ml	Serum tests
450470	Z Serum	red	1ml	Serum tests
450473	Z Serum Sep	gold	0.5ml	Serum tests
450472	Z Serum Sep	gold	0.8ml	Serum tests
450478	LH Lithium Heparin	green	0.5ml	Plasma tests
450477	LH Lithium Heparin	green	1ml	Plasma tests
450479	LH Lithium Heparin Sep	light green	0.8ml	Plasma tests
450476	K3E K3EDTA	lavender	0.25ml	Haematology
450475	K3E K3EDTA	lavender	0.5ml	Haematology
450474	K3E K3EDTA	lavender	1ml	Haematology
450484	K2E K2EDTA, transparent label	lavender	0,25 ml	Haematology
450480	K2E K2EDTA	lavender	0.5ml	Haematology
450481	FX Sod. Fluoride/Potassium Oxalate	grey	0.25ml	Glucose tests
450482	FX Sod. Fluoride/Potassium Oxalate	grey	0.5ml	Glucose tests

All tubes are packed: 100 pieces/rack; 1000 pieces/carton.

Item No.	Description	Packaging	
		Inner	Outer
MiniCollect® Accessories			
450419	Carrier Tube, amber colour, 13x75mm	100 pcs.	1000 pcs.
450420	Carrier Tube, 13x75mm	100 pcs.	1000 pcs.
450421	Funnel	100 pcs.	1000 pcs.
450430	Capillary for Serum, 250µl	100 pcs.	1000 pcs.
450431	Capillary for Serum, 80µl	100 pcs.	1000 pcs.
450437	Capillary for NaF, 250µl	100 pcs.	1000 pcs.
450432	Capillary for NaF, 80µl	100 pcs.	1000 pcs.
450433	Capillary for EDTA, 250µl	100 pcs.	1000 pcs.
450434	Capillary for EDTA, 80µl	100 pcs.	1000 pcs.
450435	Capillary for Heparin, 250µl	100 pcs.	1000 pcs.
450436	Capillary for Heparin, 80µl	100 pcs.	1000 pcs.
450427	Safety Lancet, pink, penetration depth 1 mm, blade (sterile)	200 pcs.	2000 pcs.
450428	Safety Lancet, green, penetration depth 1.5 mm, blade (sterile)	200 pcs.	2000 pcs.
450429	Safety Lancet, blue, penetration depth 2 mm, blade (sterile)	200 pcs.	2000 pcs.
450438	Safety Lancet, lavender, 28G, penetration depth 1.25 mm, needle (sterile)	200 pcs.	2000 pcs.
450439	Safety Lancet, orange, 23G, penetration depth 2.25 mm, needle (sterile)	200 pcs.	2000 pcs.

Your Power for Health


greiner bio-one

Your Power for Health


greiner bio-one



For further information, please visit our website www.gbo.com/preanalytics or contact us:

Austria (Main Office)
Greiner Bio-One GmbH
Phone (+43) 75 83 67 91-0
Fax (+43) 75 83 63 18
E-Mail office@at.gbo.com

France
Greiner Bio-One SAS
Phone (+33) 1 69 86 25 50
Fax (+33) 1 69 86 25 36
E-Mail office@fr.gbo.com

Hungary
Greiner Bio-One Hungary Kft.
Phone (+36) 96 21 30 88
Fax (+36) 96 21 31 98
E-Mail office@hu.gbo.com

Thailand
Greiner Bio-One Thailand Ltd
Phone (+66) 3 84 65 63-30
Fax (+66) 3 84 65 63-6
E-Mail office@th.gbo.com

Brazil
Greiner Bio-One Brasil
Phone (+55) 19 34 68 96 00
Fax (+55) 19 34 68 96 21
E-Mail office@br.gbo.com

Germany
Greiner Bio-One GmbH
Phone (+49) 2 01 86 18 6-0
Fax (+49) 2 01 86 18 61-2
E-Mail office@de.gbo.com

Spain
VACUETTE Espana S.A.
Phone (+34) 91 652 77 07
Fax (+34) 91 652 33 35
E-Mail info@vacuette.es

The Netherlands
Greiner Bio-One B.V.
Phone (+31) 1 72 42 09 00
Fax (+31) 1 72 44 38 01
E-Mail info@nl.gbo.com

Egypt
Greiner Bio-One Middle East
Phone (+20) 26 21 87 06
Fax (+20) 26 21 87 08
E-Mail hisham.ouda@gbo.com

Great Britain
Greiner Bio-One Ltd.
Phone (+44) 14 53 82 52 55
Fax (+44) 14 53 82 62 66
E-Mail info@uk.gbo.com

Switzerland
Greiner Bio-One VACUETTE
Schweiz GmbH
Phone (+41) 7 12 28 55 22
Fax (+41) 7 12 28 55 21
E-Mail office@ch.gbo.com

USA
Greiner Bio-One North America Inc.
Phone (+1) 70 42 61 78 00
Fax (+1) 70 42 61 78 99
E-Mail office@us.gbo.com



MiniCollect® Capillary Blood Collection System

Gentle & comfortable
for our little patients

MiniCollect® Capillary Blood Collection System



Safety Lancets

- Five types for various penetration depths and applications. Blades: pink 1mm, green 1.5mm, blue 2.0mm. Needles: lavender 1.25mm, orange 2.25mm
- The constant puncture depth of each type ensures easy and standardised collection.
- After use the blade is securely enclosed in the plastic casing.
- Controllable alignment of blade / needle during penetration.



Flexible collection with funnel or capillary

- Collection with specially adapted funnel.
- Alternatively with coated plastic capillaries.



Unique "Cross-Cuts" Caps

- Make the filling and collection of blood samples easy and hygienic.
- The cap always closes automatically.



Carrier Tube for Centrifugation

- Transparent Carrier Tube for centrifugation.
- Amber Carrier Tube for light sensitive parameters.
- Especially suitable for labelling with barcode.

MiniCollect® Tube

- The inner wall of the plastic tube is with spray-dried additive (except Sodium Citrate).
- Label with writing space.



* only suitable for venous blood, not spray-dried



MiniCollect® Capillary Blood Collection System

The Advantages at a Glance:

- Flexible rubber cap with "Cross-Cuts" means that it is not necessary to open the tube
- Maximum safety with closed system
- Label for identifying the tube
- Exact fill level determination with graduation mark on tube (1ml) or on the label (<1ml)
- For use with funnel or capillary
- Special tube design with skirt

Instructions for use:



- Cleanse the puncture area.
- Remove the lancet protective cover with pull action.

- Cleanse the puncture area.
- Remove the lancet needle cover with twist/pull action.

- Hold the puncture finger between the thumb, index and middle fingers.
- Position the lancet and press in the button to perform the puncture.
- Wipe away first blood droplet.
- By increasing and reducing the thumb pressure, blood drops form around the puncture site.

Collection with funnel

- Hold the edge of the funnel, which has been inserted through the "Cross-Cuts", to the blood droplet to transfer blood into the tube.

Collection with capillary

- Hold the capillary and tube horizontally.
- Hold the tip of the capillary to the blood droplet. The tube will fill automatically due to the capillary effect.
- Sequence of collection:
 - 1) EDTA tubes
 - 2) Tubes with other additives
 - 3) Tubes without additives

- After blood collection, gently remove the funnel or capillary from the tube and dispose of. The cap closes automatically.
- After collection, invert the tube gently several times.